



valid until: December 16, 2027

# Fraunhofer

## TESTED<sup>®</sup> DEVICE

INFICON  
UL6000 Fab Plus  
Report No. IN 2211-1364

### Single product Particle Emission

DUPLICATE

### Qualification Certificate

This is to certify that the product mentioned above, provided by

**INFICON GmbH**  
Cologne, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number IN 2211-1364.

The Mobile vacuum leak tester - UL6000 Fab Plus (color: light gray/blue) was assessed in compliance with ISO 14644-14. When operated under the specified test conditions, it is suitable for use in cleanrooms fulfilling the specifications of the following Air Cleanliness Classes according to ISO 14644-1:

Test parameter(s)	Air Cleanliness Class
Ultra: Pump Hipace80 (TMP) = 1500Hz; Pump Neody36G (forepump) = 175 Hz	5
Exhaust air level	5
<b>Overall result</b>	<b>5</b>

Please note: Transport damages, incorrect installation, corrosion, oil leakage, aging behavior etc. can influence the test result.

IN 2211-1364  
Report No. first document

Stuttgart, December 16, 2022  
Place, date of first document issued

--  
Report No. current document

--  
Place, current date

on behalf of 

This document only applies to the named product in its original state and is valid for a period of 5 years from the date the first document was issued. The document can be verified under [www.tested-device.com](http://www.tested-device.com). Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.