



# Fraunhofer

## TESTED<sup>®</sup> DEVICE

M. Braun Inertgas-Systeme GmbH  
Flow Box (Project 11144)  
**Report No. MB 1703-892**

### Cleanroom Acceptance

### Qualification Certificate

This is to certify that the product mentioned above, provided by

**M. Braun Inertgas-Systeme GmbH**  
Garching, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE  
bearing the report number MB 1703-892

In the state "at rest", the MB Laminar Flow Box (Project 11144)  
fulfills the following requirements

Function test	Result
Air Cleanliness Class according to ISO 14644-1	ISO Class 3

DUPLICATE

The validity of this certificate applies only to the mentioned product in this particular condition for a duration of 5 years.

Further information:

[www.tested-device.com](http://www.tested-device.com)

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

Stuttgart, March 20, 2017

Place, date of first document issued

--

Place, current date

on behalf of

  
Frank Bürger, Project manager Fraunhofer IPA