



valid until: July 28, 2028

# Fraunhofer

## TESTED<sup>®</sup> DEVICE

Bimos - Interstuhl  
Labsit 9123-2000-3285  
**Report No. IN 2307-1438**

### Single product Particle Emission

### Qualification Certificate

This is to certify that the product mentioned above, provided by

**Interstuhl Büromöbel GmbH & Co. KG**  
Tieringen, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE  
bearing the report number IN 2307-1438.

The chair Bimos - Labsit 9123-2000-3285-521304-702 (color:  
black/anthrazit) was assessed in compliance with ISO 14644-14.  
When operated under the specified test conditions, it is suitable  
for use in cleanrooms fulfilling the specifications of the following  
Air Cleanliness Classes according to ISO 14644-1:

Test parameter(s)	Air Cleanliness Class
Seat (F = 1200 N; 12 cycles/min)	4
Backseat (F = 350 N; 12 cycles/min)	2
<b>Overall result</b>	<b>4</b>

Please note: Transport damages, incorrect installation, oil leakage  
aging behavior, corrosion etc. can influence the test result.

This document only  
applies to the named  
product in its original state  
and is valid for a period of  
5 years from the current  
date the document was  
issued. The document can  
be verified under  
**www.tested-device.com**.

Detailed information and  
parameters of the test  
environment can be found  
in the Fraunhofer IPA test  
report.

IN 2307-1438  
Report No. first document

Stuttgart, March 25, 2017  
Place, date of first document issued

IN 2307-1438  
Report No. current document

Stuttgart, 28. Juli 2023  
Place, current date

on behalf of   
Dr.-Ing. Frank Bürger, Project Manager Fraunhofer IPA