



Fraunhofer

TESTED[®] DEVICE

Dr. Staiger, Mohilo
Type 4740A
Report No. KI 1105-552

Qualification Certificate

This is to certify that the product mentioned above, provided by

Dr. Staiger, Mohilo + Co. GmbH

A Kistler Group Company
Lorch, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE
bearing the report number KI 1105-552.

The force/displacement measuring system Type 4740A is suitable
for use in cleanrooms fulfilling the Air Cleanliness Class 1
according to ISO 14644-1.

Detailed information and parameters of the test environment can
be found in the Fraunhofer IPA test report.

Stuttgart, June 30, 2011


i. A. _____
Project manager

The validity of this certificate
applies only to the mentioned
product in this particular
condition for a duration of
5 years.

Further information:
www.tested-device.com

DUPLICATE