



Fraunhofer

TESTED[®] DEVICE

OCTANORM
Al-composite-panels C826
Report No. OC 1005-524

DUPLICATE

Qualification Certificate

This is to certify that the product mentioned below provided by the company

OCTANORM-Vertriebs-GmbH für Bauelemente

Raiffeisenstraße 39
70794 Filderstadt
Germany

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number OC 1005-524.

The sample is resistant to: Formalin (37 %), Isopropanol, Ammonia (25 %), Hydrogen peroxide (30 %) and Phosphoric acid (30 %).

The sample is partially resistant to: Sodium hypochloride (5 %) and Sulphuric acid (5 %).

The sample is not resistant to: Caustic soda (5 %), Hydrochloric acid (5 %) and Peracetic acid (15 %).

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

The validity of this certificate applies only to the above mentioned product in this particular condition. For further information visit our website at www.ipa-qualification.com.

Stuttgart, September 1, 2010

i.A. 
Project manager