



# Fraunhofer

## TESTED<sup>®</sup> DEVICE

Jung Gummitechnik GmbH  
Jugitec Pharma Plus  
Report No. JU 1504-756

### Qualification Certificate

This is to certify that the product mentioned above, provided by

**Jung Gummitechnik GmbH**  
Einhausen, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number JU 1504-756.

Chemical resistance	1 h	3 h	6 h	24 h
Formalin 37 %	0	0	0	0
Ammoniac 25 %	0	0	0	0
Hydrogen peroxide 30 %	0	0	0	0
Sulphuric acid 5 %	0	0	0	0
Phosphoric acid 30 %	0	0	0	0
Peracetic acid 15 %	1	2	2	2
Hydrochloric acid 5 %	0	0	0	0
Isopropanol 100 %	1	2	2	4
Sodium hydroxide 5 %	0	0	0	0
Sodium hypochlorite 5 %	0	0	0	0
<b>Classification</b>	<b>1/very good</b>			

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

Stuttgart, May 26, 2015

i.A.   
Frank Bürger, Project manager Fraunhofer IPA

DUPLICATE

The validity of this certificate applies only to the mentioned product in this particular condition for a duration of 5 years.

Further information:  
[www.tested-device.com](http://www.tested-device.com)