



Fraunhofer

TESTED[®] DEVICE

Roche Diagnostics GmbH
MP2L stoneware tile
Report No. RO 1608-852

Chemical Resistance

Qualification Certificate

This is to certify that the product mentioned above, provided by

Roche Diagnostics GmbH
Mannheim, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number RO 1608-852.

The chemical resistance of MP2L stoneware tile was classified according to ISO 4628-1 and VDI 2083-17 with the following result:

Chemical resistance	Classification
Purified water 100 %	excellent
Ethanol 100 %	excellent

The validity of this certificate applies only to the mentioned product in this particular condition for a duration of 5 years.

Further information:

www.tested-device.com

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

Stuttgart, November 30, 2016

on behalf of 
Frank Bürger, Project manager Fraunhofer IPA