



# Fraunhofer

## TESTED<sup>®</sup> DEVICE

KUKA Roboter GmbH  
LBR iiwa 7 R800 CR

**Report No. KU 1707-925**

DUPLICATE

Statement of  
Qualification

Particle Emission

# Statement of Qualification

**Customer**  
 KUKA Roboter GmbH  
 Zugspitzstrasse 140  
 86165 Augsburg  
 Germany

**Component tested**

Category: Automation Components

Subcategory: Robotics

Product name: LBR iiwa 7 R800 CR  
 (manufacturing date: 5/2017; article number: 10027613 ; serial number: 982698; payload: 7 kg; reach: 800 mm)

## Random sampling of particle emissions (airborne) at representative sites

Standards/Guidelines: ISO 14644-1, -14  
 The norms stated generally refer to the version valid at the time of the tests.

Test devices: Optical particle counter:  
 LasAir II 110 and LasAir III 110 with measuring ranges  $\geq 0.1 \mu\text{m}$ ,  $\geq 0.2 \mu\text{m}$ ,  $\geq 0.3 \mu\text{m}$ ,  $\geq 0.5 \mu\text{m}$ ,  $\geq 1.0 \mu\text{m}$  and  $> 5.0 \mu\text{m}$

Test environment parameters:

- Cleanroom Air Cleanliness Class (according to ISO 14644-1):..... ISO 1
- Airflow velocity:.....0.45 m/s
- Airflow pattern:..... vertical laminar flow
- Temperature: .....22°C  $\pm$  0.5°C
- Relative humidity: ..... 45 %  $\pm$  5 %

Test procedure parameters:

- Capacity: ..... 40 % and 80 % of maximum utilization
- Attached payload: ..... 7 kg
- Pause between cycles: .....0s
- Operation of each axis:..... separately
- Movement of each axis:
  - Axis 1: .....-170° until 170°
  - Axis 2: .....-90° until 90°
  - Axis 3: .....-90° until 90°
  - Axis 4: .....-120° until 120°
  - Axis 5: .....-170° until 170°
  - Axis 6: .....-120° until 120°
  - Axis 7: .....-175° until 175°

## Test result / Classification

When operated under the specified test conditions, the robot LBR iiwa 7 R800 CR is suitable for use in cleanrooms fulfilling the specifications of the following Air Cleanliness Classes according to ISO 14644-1:

Test parameter(s)	Air Cleanliness Class
Capacity = 40 %	2
Capacity = 80 %	3
<b>Overall result</b>	<b>3</b>

The measuring devices used for the qualification tests are calibrated at regular intervals; their results can be traced back to national and international standards. In cases where no national standards exist, the test procedure implemented complies with the technical regulations and norms applicable at the time of the test. The relevant documentation can be viewed on request at any time.

For further information about the test environment and parameters, please refer to the Fraunhofer IPA test report.

Fraunhofer Institute for Manufacturing Engineering and Automation IPA

Department of Ultraclean Technology and Micromanufacturing

Nobelstrasse 12  
 70569 Stuttgart  
 Germany

KU 1707-925  
 Report No. first document

Stuttgart, October 12, 2017  
 Place, date of first document issued

--  
 Report No. current document

--  
 Place, current date

on behalf of   
 Dr.-Ing. Frank Bürger, Project Manager Fraunhofer IPA