



# Fraunhofer

## TESTED<sup>®</sup> DEVICE

Rockwool B.V., Rockfon  
Medicare Block (prototype)  
**Report No. RO 1212-627**

DUPLICATE

Statement of  
Qualification

# Statement of Qualification

**Customer:** Rockwool B.V., Rockfon  
Industrieweg 15  
6045 JG Roemond  
The Netherlands

**Component tested:**

Category: Cleanroom Facilities  
Subcategory: Wall / Ceiling / Floor  
Type: Ceiling system Medicare Block (prototype)

**Test results / Classification:**  
(according to ISO 14644-1)

The Medicare Block (prototype) ceiling system is suitable for use in cleanrooms fulfilling the Air Cleanliness Class 2.

It must be pointed out, that according to ISO 14644-1 cleanrooms classes 1 to 5 have a high number of filters, which makes the use of ceiling elements partly impossible. Cleanrooms with horizontal laminar flow are an exception.

**Random check measurements of particle emission (airborne) at representative points**

Test procedure: According to VDI 2083-9.1; ISO 14644-1  
Each standard states refers to the version valid at the time of testing.

Measuring instruments: Optical Particle Counter:  
Model LasAir II 110 manufactured by PMS with measuring channels of  
0.1 µm, ≥ 0.2 µm, ≥ 0.3 µm, ≥ 0.5 µm, ≥ 1.0 µm and ≥ 5.0 µm

Test parameters of the test environment:

- Cleanroom Air Cleanliness Class (according to ISO 14644-1):..... ISO 1
- Air flow velocity:.....0.45 m/s
- Air flow guidance: .....vertical unidirectional air flow
- Temperature: .....22 °C ± 0.5 °C (71.6 °F ± 0.9 °F)
- Relative humidity: ..... 45 % ± 5 %

Test parameters of the test execution:

The ceiling system was tested in the following grid system:

- T-grid system: ..... Chicago Metallic Continental
- Serial number:..... 1267081369
- Width: ..... 24 mm
- Height:..... 38 mm

The ceiling system was stressed as follows:

- Impact sound: .....5 Hz to 50 Hz
- Average oscillation velocity:..... v = 0.1495 mm/s
- Average oscillation acceleration:..... a = 0.083 m/s<sup>2</sup>
- Average oscillation of the system: .....s = 0.00047 mm

DUPLICATE

DUPLICATE

The measuring equipment used for the qualification is regularly calibrated and is based on national and international standards. In the case where no national standards exist, the measuring procedure used corresponds with technical regulations and norms valid at the time of the measurement. The documents drawn up for this procedure are available for viewing.

The validity of this certificate applies only to the mentioned product in this particular condition for a duration of 5 years.  
Further information: [www.tested-device.com](http://www.tested-device.com).

Fraunhofer Institute for  
Manufacturing Engineering and Automation IPA

Department Ultraclean Technology  
and Micromanufacturing

Nobelstrasse 12  
70569 Stuttgart  
Germany

Stuttgart, March 8, 2013  
Place, Date

  
i. A.   
Project manager