



Fraunhofer

**TESTED[®]
DEVICE**

Jung Rubbertec
Jugitec ISOflex Plus
Report No. JU 1511-794

DUPLICATE

Statement of
Qualification

Hydrogen Peroxide
Absorption/Desorption

Statement of Qualification

Customer
 Jung Gummitechnik GmbH
 Robert-Bosch-Strasse 12
 64683 Einhausen
 Germany

Component tested
 Category: Materials
 Subcategory: Consumables
 Product name: Jugitec ISOflex Plus
 (manufacturing date: 6/2015; color: translucent;
 charge number: prototype)

Hydrogen peroxide absorption / desorption
 Methodics:

- Material exposure to vaporized hydrogen peroxide for a defined duration using an emission test cell
- Aeration (with ambient air) of the test setup with continuous monitoring of the decreasing hydrogen peroxide concentration
- Calculation of the k-value as time needed to reach 1/10 of the maximum hydrogen peroxide concentration measured at start of the aeration

 Air-conditioned laboratory environment: Temperature:22°C ± 0.5°C
 Test procedure parameters:

- Emission test cell volume: 16.5 cm³
- Exposed surface area: 33 cm²
- H₂O₂ vapor concentration: 50 ± 20 ppm (V)
- Exposure duration: 60 min
- Air exchange rate during aeration: 100 min⁻¹

Test result / Classification

The hydrogen peroxide absorption/desorption of the glove Jugitec ISOflex Plus was investigated at the stated test parameters, providing the following test result:

Ø k-value [min]	Standard deviation [min]	Classification
1	0.0	non-absorptive

The k-value represents the required decay time to reduce the hydrogen peroxide vapor concentration measured at the beginning of the aeration phase to one tenth of the original concentration. The material classification is based on three separate measurements. The blank value of the test setup is subtracted from each measurement value. The average k-value is transferred to the following classification:

- < 5 min: non-absorptive
- 5-15 min: fast
- 15-60 min: average
- > 60 min: slow
- Not determinable: catalytic activity

The k-value can only be used to make a comparative material assessment. Provided the maximum hydrogen peroxide vapor concentration during material exposure is within the defined limit, it does not affect the calculated k-values.

The measuring devices used for the qualification tests are calibrated at regular intervals; their results can be traced back to national and international standards. In cases where no national standards exist, the test procedure implemented complies with the technical regulations and norms applicable at the time of the test. The relevant documentation can be viewed on request at any time.

For further information about the test environment and parameters, please refer to the Fraunhofer IPA test report.

Fraunhofer Institute for Manufacturing Engineering and Automation IPA

Department of Ultraclean Technology and Micromanufacturing

Nobelstrasse 12
 70569 Stuttgart
 Germany

Stuttgart, December 11, 2015

Place, date of first document issued

--

Place, current date

on behalf of 
 Frank Bürger, Project Manager Fraunhofer IPA