



# Fraunhofer

## TESTED<sup>®</sup> DEVICE

B Medical Systems  
PRECISION LINE - PF260  
**Report No. BM 2202-1300**

Single product  
Particle Emission

### Qualification Certificate

This is to certify that the product mentioned above, provided by

**B Medical Systems S.à r.l.**  
Hosingen, Luxembourg

has been awarded a Fraunhofer certificate TESTED DEVICE  
bearing the report number BM 2202-1300.

The PRECISION LINE - COMBINED PHARMACY REFRIGERATOR  
AND LABORATORY FREEZER PF260 (color: light gray) was assessed  
in compliance with ISO 14644-14. When operated under the  
specified test conditions, it is suitable for use in cleanrooms  
fulfilling the specifications of the following Air Cleanliness Class  
according to ISO 14644-1:

Test parameter(s)	Air Cleanliness Class
Temperature = 5°C; -32°C	5
Overall result	

The material used for the insulating pipe is not suitable for clean-  
rooms. The material carries a risk of contamination. The PRECISION  
LINE - COMBINED PHARMACY REFRIGERATOR AND LABORATORY  
FREEZER PF260 must be used without the insulating pipe.

Please note: Transport damages, incorrect installation, oil leakage,  
aging behavior, corrosion etc. can influence the test result.

BM 2202-1300  
Report No. first document

Stuttgart, June 13, 2022  
Place, date of first document issued

--  
Report No. current document

--  
Place, current date

on behalf of   
Dr.-Ing. Frank Bürger, Project Manager Fraunhofer IPA

This document only  
applies to the named  
product in its original state  
and is valid for a period of  
5 years from the date the  
first document was issued.  
The document can be  
verified under  
**[www.tested-device.com](http://www.tested-device.com)**.

Detailed information and  
parameters of the test  
environment can be found  
in the Fraunhofer IPA test  
report.