





Fraunhofer TESTED® DEVICE Coroplast Coroflex Food Hygienic 3300 Report No. CO 1509-784

Statement of Qualification

Biological Resistance

Statement of Qualification

Customer

Category:

Subcategory

Product name:

Biological resistance test

Test environment parameters:

Test procedure parameters:

Standards/Guidelines:

Component tested

Coroplast Fritz Müller GmbH & Co. KG Wittener Strasse 271 42279 Wuppertal Germany

Materials

Plastics

Coroflex Food Hygienic 3300

ISO 846; VDI 2083-18

- Aspergillus niger

– Penicillium funiculosum

Pseudomonas aeruginosa

– Paecilomyces variotii

Microbiological laboratory:.....

(manufacturing date: 8/2015; color: light gray; serial number: 29-3300)

The norms stated generally refer to the version valid at the time of the tests.

- Gliocladium virens

- Chaetomium globosum

• Fungus test (Procedure A) using spore suspension containing:

• Bacteria test (Procedure C) using bacteria suspension containing

95 % relative humidity; visual analysis after four (4) weeks

• Incubation at 24 °C (Procedure A) respectively 29 °C (Procedure C) and

....S2

Test result/Classification

The biological resistance of the cable sheathing material Coroflex Food Hygienic 3300 with regard to growth intensity was investigated in accordance with ISO 846 and classified with the following result:

Biological resistance	Growth intensity	Classification
Fungi (Procedure A)	0	excellent
Bacteria (Procedure C)	0	excellent
Overall result	0	excellent

The classification is based on a worst-case consideration of Procedures A and C. In the process, growth intensity was assessed according to the classification system used in ISO 846:

0 = excellent1 = very good2 = good

The measuring devices used for the qualification tests are calibrated at regular intervals; their results can be traced back to national and international standards. In cases where no national standards exist, the test procedure implemented complies with the technical regulations and norms applicable at the time of the test. The relevant documentation can be viewed on request at any time.

For further information about the test environment and parameters, please refer to the Fraunhofer IPA test report.

Fraunhofer Institute for Manufacturing Engineering and Automation IPA

Stuttgart, March 20, 2016 Place, date of first document issued

Department of Ultraclean Technology and Micromanufacturing

Nobelstrasse 12 70569 Stuttgart Germany

Place, current date

on behalf of R. Bri Frank Bürger, Project Manager Fraunhofer IPA



3 = weak 4 = very weak5 = none

This document only applies to the named product in an unchanged state and is valid from the date of issue for a period of 5 years. The document can be verified under www.tested-device.com.