



Fraunhofer

TESTED[®] DEVICE

CIM med GmbH
CIM support Arm
Report No. CI 1310-672

Qualification Certificate

This is to certify that the product mentioned above, provided by

CIM med GmbH
München, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number CI 1310-672.

The CIM support Arm can be simply wiped clean using ultra-pure water. However, the fluorescence test identified several critical areas. These areas need to be cleaned especially thoroughly, or a more complex procedure needs to be implemented, for example dismantling some components before cleaning. It is very difficult to clean these areas of the support arm effectively. Contamination could collect as a result. By selecting appropriate materials and avoiding as many gaps, open depressions and undercuts as possible, the cleanability of the support arm can be improved significantly.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

Stuttgart, November 27, 2013

i.A. 
Project manager

DUPLICATE

The validity of this certificate applies only to the mentioned product in this particular condition for a duration of 5 years.

Further information:
www.tested-device.com