



# Fraunhofer

## TESTED<sup>®</sup> DEVICE

Skan AG  
Glove CSM / Tosoh 6/10 mm  
**Report No. SK 1503-753**

### Qualification Certificate

This is to certify that the product mentioned above, provided by

**Skan AG**  
Allschwil, Switzerland

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number SK 1503-753.

Ø k-value [min]	Standard deviation [min]	Classification
0.0	0.0	non-absorptive

The k-value (expressed in minutes) represents the required decay time to reduce the hydrogen peroxide vapor concentration measured at the beginning of the aeration phase to one tenth of the original concentration. The material classification is based on three separate measurements. The blank value of the test setup is subtracted from each measurement value.

- <5 min: ..... non-absorptive
- 5-15 min: ..... fast
- 15-60 min: ..... average
- >60 min: ..... slow
- Not determinable: ..... catalytic activity

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

Stuttgart, May 4, 2015

i.A.   
Udo Gommel, Project manager Fraunhofer IPA

DUPLICATE

The validity of this certificate applies only to the mentioned product in this particular condition for a duration of 5 years.

Further information:  
[www.tested-device.com](http://www.tested-device.com)