



# Fraunhofer

## TESTED<sup>®</sup> DEVICE

Jung Rubbertec  
Jugitec ISOflex Plus  
Report No. JU 1511-794

### Hydrogen Peroxide Absorption/Desorption

### Qualification Certificate

This is to certify that the product mentioned above, provided by

**Jung Rubbertec**  
Einhausen, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number JU 1511-794.

The hydrogen peroxide absorption/desorption of the glove Jugitec ISOflex Plus (color: translucent) was investigated at the stated test parameters, providing the following test result:

Ø k-value [min]	Standard deviation [min]	Classification
1	0.0	non-absorptive

The k-value represents the required decay time to reduce the hydrogen peroxide vapor concentration measured at the beginning of the aeration phase to one tenth of the original concentration. The material classification is based on three separate measurements. The blank value of the test setup is subtracted from each measurement value.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

Stuttgart, December 11, 2015

on behalf of   
Frank Bürger, Project manager Fraunhofer IPA

The validity of this certificate applies only to the mentioned product in this particular condition for a duration of 5 years.  
Further information:  
[www.tested-device.com](http://www.tested-device.com)

DUPLICATE