



Fraunhofer

TESTED[®] DEVICE

Roche Diagnostics GmbH
MP2L joint seal

Report No. RO 1608-841

Chemical Resistance

Qualification Certificate

This is to certify that the product mentioned above, provided by

Roche Diagnostics GmbH
Mannheim, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE
bearing the report number RO 1608-841.

The chemical resistance of the MP2L joint seal was classified
according to ISO 4628-1 and VDI 2083-17 with the following
result:

Chemical resistance	Classification
Purified water 100 %	excellent
Ethanol 100 %	weak

The validity of this certificate
applies only to the mentioned
product in this particular
condition for a duration of
5 years.

Further information:

www.tested-device.com

DUPLICATE

Detailed information and parameters of the test environment can
be found in the Fraunhofer IPA test report.

Stuttgart, November 30, 2016

on behalf of 
Frank Bürger, Project manager Fraunhofer IPA