



Cleanroom[®] Suitable Materials

Roche Diagnostics GmbH
Report No. RO 1608-841

MP2L joint seal
Hydrogen Peroxide Absorption /
Desorption: fast

FLOORING & COATING

Hydrogen Peroxide Absorption/Desorption

Qualification Certificate

We hereby certify that the material stated above, provided by

Roche Diagnostics GmbH
Mannheim, Germany

has been awarded the Fraunhofer IPA CSM Certificate of
Qualification with the report number RO 1608-841.

The hydrogen peroxide absorption/desorption of the MP2L joint
seal was investigated at the stated test parameters, providing the
following test result:

Ø k-value [min]	Standard deviation [min]	Classification
9.5	2.3	fast

- <5 min: non-absorptive
- 5-15 min: fast
- 15-60 min: average
- >60 min: slow
- Not determinable: catalytic activity

Detailed information regarding specific product data such as date of
manufacture, color, and serial number as well as test environment
and parameters can be obtained from the CSM test report issued by
the Fraunhofer-Gesellschaft.

Stuttgart, November 30, 2016

on behalf of 
Udo Gommel, Project manager Fraunhofer IPA

DUPLICATE

The validity of this
certificate applies only to
the mentioned product in
this particular condition for
a duration of
5 years.
Further information:
www.tested-device.com.
Please visit www.ipa-csm.com.