



Cleanroom[®] Suitable Materials

Roche Diagnostics GmbH
Report No. RO 1608-853

PQ18 granite
Particle (vs. PA6): ISO 5

FLOORING & COATING

Particle Emission

Qualification Certificate

We hereby certify that the material stated above, provided by

Roche Diagnostics GmbH
Mannheim, Germany

has been awarded the Fraunhofer IPA CSM Certificate of Qualification with the report number RO 1608-853.

The level of particulate contamination of the PQ18 granite emitted during application of tribological stress on the material pairing specified lies within the permissible values of the corresponding Air Cleanliness Classes in accordance with ISO 14644-1:

Material pairing and test		Suitable for Air Cleanliness Class (ISO 14644-1)
PQ18 granite	vs. PA6	ISO Class: 5
Reel-on-disc test Normal force 300N		

DUPLICATE

Detailed information regarding specific product data such as date of manufacture, color, and serial number as well as test environment and parameters can be obtained from the CSM test report issued by the Fraunhofer-Gesellschaft.

Stuttgart, November 30, 2016

on behalf of 
Udo Gommel, Project manager Fraunhofer IPA

The validity of this certificate applies only to the mentioned product in this particular condition for a duration of 5 years.

Further information:
www.tested-device.com.
Please visit www.ipa-csm.com.