



valid until: November 16, 2028

Fraunhofer

TESTED[®] DEVICE

SG Armaturen AS
KENT Rotabond 2000
Report No. **SG 2303-1398**

Single product Chemical Resistance

Qualification Certificate

This is to certify that the product mentioned above, provided by

SG Produktion A/S
Odense N, Denmark

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number SG 2303-1398.

The chemical resistance of KENT Rotabond 2000 (color: white) was classified according to ISO 4628-1 and VDI 2083 Part 17 with the following result:

Chemical resistance (24 h)			
Formalin 37 %	excellent	Peracetic acid 15 %	none
Ammoniac 25 %	weak	Hydrochloric acid 5 %	good
Hydrogen peroxide 30 %	none	Isopropanol 100 %	excellent
Sodium hypochlorite 5 %	weak	Sodium hydroxide 5 %	weak
Phosphoric acid 30 %	good	Sulfuric acid 5 %	very weak

DUPLICATED

SG 2303-1398
Report No. first document

Stuttgart, November 16, 2023
Place, date of first document issued

--
Report No. current document

--
Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, Project Manager Fraunhofer IPA

This document only applies to the named product in its original state and is valid for a period of 5 years from the date the first document was issued. The document can be verified under www.tested-device.com.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.